

# virotype<sup>®</sup> Mix +IC(JOE)-DNA Product Sheet

For real-time amplification of bacterial and viral DNA and Internal Control DNA

---

**REF** 96 reactions (cat. no. MX289975)

**REF** 480 reactions (cat. no. MX289977)



INDICAL BIOSCIENCE GmbH, Deutscher Platz 5b,  
04103 Leipzig, Germany

# Contents

<b>virotype Mix +IC(JOE)-DNA</b>	<b>(96)</b>	<b>(480)</b>
<b>Cat. no.</b>	<b>MX289975</b>	<b>MX289977</b>
<b>Number of reactions</b>	<b>96</b>	<b>480</b>
virotype Mix +IC(JOE)-DNA		
Mix includes buffer, enzymes, primers and probe for Internal Control and Internal Control DNA	2 x 900 µl	10 x 900 µl

## Intended use

The virotype Mix +IC(JOE)-DNA contains synthetically derived nucleic acids and enables real-time amplification of bacterial and viral DNA and the Internal Control DNA.

**For molecular biology applications. This product is not intended for the diagnosis, prevention or treatment of a disease.**

## Quality control

Each lot of this product was manufactured and released in accordance with INDICAL's ISO-certified Quality Management System.

## Storage and handling

Upon receipt, store the material at  $-30^{\circ}\text{C}$  to  $-15^{\circ}\text{C}$ , away from any sources of contaminating DNA or RNA, especially amplified DNA products and material used for PCR master mixes. Protect from light during storage and handling. Use aerosol barrier pipet tips for pipetting. Do not use after the expiration date printed on the label. Avoid repeated thawing and freezing ( $>2\times$ ), because this may reduce performance. If used intermittently, aliquot the material and freeze.

## Safety information

When working with chemicals, always wear a suitable lab coat, disposable gloves and protective goggles. For more information, please consult the appropriate safety data sheets (SDSs). These are available from your regional sales manager or at **[compliance@indical.com](mailto:compliance@indical.com)**.

All sample residues and objects that have come into contact with samples must be decontaminated or disposed of as potentially infectious material.

## Warning and precautions

Dispose of all materials in accordance with national and local regulations.

# Principle and procedure

The virotype Mix +IC(JOE)-DNA provides multiplex real-time PCR detection of one or more nucleic acid targets and the heterologous Internal Control using dual-labeled probes.

The Internal Control DNA is already added to the virotype Mix +IC(JOE)-DNA and will result in positive signals in No Template Controls (NTC) reactions. Missing Internal Control signals in the sample reaction may indicate the presence of inhibitors or other issues with the sample reaction.

## Things to do before starting

- Thaw the virotype Mix +IC(JOE)-DNA on ice and protected from light.
- Before use, spin the mix briefly.

## Preparation of the PCR reaction

- It is recommended to use 18 µl of the virotype Mix +IC(JOE)-DNA per 25 µl reaction.

## Filter settings

- The probe detecting the Internal Control is JOE™ labeled.
- The virotype Mix +IC(JOE)-DNA contains a low ROX™ dye concentration as a passive reference dye. This allows the use of instruments like Applied Biosystems™ 7500, ViiA™ 7 and QuantStudio Systems.

## Cycling conditions

<b>Step</b>	<b>Temperature</b>	<b>Time</b>	<b>Number of cycles</b>
<b>Initial activation</b>	95°C	2 min	1
<b>2-step cycling</b>			
Denaturation	95°C	5 s	40
Annealing/extension	60°C*	30 s	

\* Fluorescence data collection

# Notes

# Notes

**Trademarks:** virotype® (INDICAL BIOSCIENCE GmbH). Applied Biosystems™, JOE™, ROX™, ViiA™ 7, QuantStudio (Life Technologies Corporation). Licensed probes manufactured by Integrated DNA Technologies, Inc. Registered names, trademarks, etc. used in this document, even when not specifically marked as such, are not to be considered unprotected by law. For updated license terms, see [www.indical.com](http://www.indical.com).

HB-2210-EN-003 © 2018-2020 INDICAL BIOSCIENCE GmbH, all rights reserved.

## Change index

<b>Product Sheet</b>	<b>Version</b>	<b>Change</b>
HB-2210-EN-003	February 2020	Product name change
HB-2210-EN-002	October 2019	Addition of new package size
HB-2210-001	May 2018	Product launch